


## Mycobacterium leprae Control Slides

### Tissue:

- (+) Human tissue with *M. leprae* infection
- (-) Human tissue

Note: Tissue type may vary due to availability.



Fite Stain

Fixative: 10% Neutral buffered formalin

Embedding: Paraffin

Thickness: 4  $\mu$ m

Mounting: Mounted on positively charged slides made of pure white glass; air dried

STORAGE: 2-8°C.

INTENDED USE: Deparaffinize and rehydrate before use.

M. leprae	purple red staining
background	Light blue

Positive tissue controls should only be utilized for monitoring the correct performance of processed tissues and test reagents, rather than as an aid in the interpretation of samples. If the positive tissue controls fail to demonstrate positive staining, results with the test specimens should be considered invalid.